



Fraunhofer

TESTED[®] DEVICE

CAITRON GmbH
Cleanroom HMI CR-series
Report No. CA 2005-1157

Product series
Hygienic Design

Qualification Certificate

This is to certify that the product mentioned above, provided by

CAITRON GmbH
Ramerberg, Germany

has been awarded a Fraunhofer certificate TESTED DEVICE bearing the report number CA 2005-1157.

The Cleanroom HMI CR-series is principally suitable for use in hygienic areas up to the following GMP Class according to EU GMP Annex 1:

Suitability
up to GMP Class A / B

However, this recommendation only pertains to the operating utility when in a resting state. An overall assessment can only be made after its installation in the production line.

This document only applies to the named product in its original state and is valid for a period of 5 years from the date the first document was issued. The document can be verified under www.tested-device.com.

Detailed information and parameters of the test environment can be found in the Fraunhofer IPA test report.

DUPLICATE

CA 2005-1157
Report No. first document

Stuttgart, July 3, 2020
Place, date of first document issued

--
Report No. current document

--
Place, current date

on behalf of 
Dr.-Ing. Frank Bürger, Project Manager Fraunhofer IPA